

Instructions for Use



Amyloid β Auto-Ab (Serum) ELISA

RUO

REF

EIA-5099



96



Legal Manufacturer:

DRG

DRG Instruments GmbH, Germany

Division of DRG International, Inc

Frauenbergstr. 18, D-35039 Marburg

Telefon: +49 (0)6421-17000 Fax: +49-(0)6421-1700 50

Internet: www.drg-diagnostics.de

E-mail: drg@drg-diagnostics.de

Distributed by:

Contents / Inhaltsverzeichnis / Contenuti / Contenido / Contenu

1	INTRODUCTION.....	2
2	PRINCIPLE OF THE TEST	2
3	WARNINGS AND PRECAUTIONS	3
4	REAGENTS	4
5	SPECIMEN COLLECTION AND PREPARATION.....	5
6	ASSAY PROCEDURE	6
7	EXPECTED NORMAL VALUES.....	7
8	QUALITY CONTROL	7
9	PERFORMANCE CHARACTERISTICS.....	8
10	LIMITATIONS OF USE	9
11	LEGAL ASPECTS	9
12	REFERENCES / LITERATURE.....	9
	SYMBOLS USED WITH DRG ELISAS	10

1 INTRODUCTION

1.1 Intended Use

The **DRG Amyloid β Auto-Ab ELISA** is an enzyme immunoassay for the quantitative *in vitro* measurement of autoantibodies against Amyloid Beta ($A\beta$) in human serum.

For research use only.

1.2 Summary and Explanation

The Amyloid β ($A\beta$) protein consists of 40-42 (43) amino acids and is cleaved from Amyloid Precursor Protein (APP) by β -secretase and subsequent γ -secretase. The increased release of $A\beta_{42}/A\beta_{43}$ leads to the abnormal deposition of human $A\beta$ and the associated neurotoxicity in the brains of affected individuals.^{1,2}

2 PRINCIPLE OF THE TEST

The DRG Amyloid β Auto-Ab ELISA Kit is a solid phase enzyme-linked immunosorbent assay (ELISA) based on the sandwich principle.

The microtiter wells are coated with (1-42) β -Amyloid. An aliquot of patient sample containing $A\beta$ -autoantibodies is incubated in the coated well with assay buffer. After washing a second incubation with enzyme conjugate, which is goat anti-human IgG conjugated to horseradish peroxidase, follows. After incubation the unbound conjugate is washed off.

The amount of bound peroxidase is proportional to the concentration of autoantibodies against $A\beta$ in the sample.

Having added the substrate solution, the intensity of colour developed is proportional to the concentration of autoantibodies against $A\beta$ in the patient sample.

3 WARNINGS AND PRECAUTIONS

1. For research use only. For professional use only.
2. All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
3. Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood.
4. The microplate contains snap-off strips. Unused wells must be stored at 2 °C to 8 °C in the sealed foil pouch and used in the frame provided.
5. Pipetting of samples and reagents must be done as quickly as possible and in the same sequence for each step.
6. Use reservoirs only for single reagents. This especially applies to the substrate reservoirs. Using a reservoir for dispensing a substrate solution that had previously been used for the conjugate solution may turn solution colored. Do not pour reagents back into vials as reagent contamination may occur.
7. Mix the contents of the microplate wells thoroughly to ensure good test results. Do not reuse microwells.
8. Do not let wells dry during assay; add reagents immediately after completing the rinsing steps.
9. Allow the reagents to reach room temperature (21-26°C) before starting the test. Temperature will affect the absorbance readings of the assay. However, values for the patient samples will not be affected.
10. Never pipet by mouth and avoid contact of reagents and specimens with skin and mucous membranes.
11. Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.
12. Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
13. Handling should be done in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
14. Do not use reagents beyond expiry date as shown on the kit labels.
15. All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiterplate readers.
16. Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
17. Avoid contact with *Stop Solution* containing 0.5 M H₂SO₄. It may cause skin irritation and burns.
18. Some reagents contain Proclin 300, BND and/or MIT as preservatives. In case of contact with eyes or skin, flush immediately with water.
19. TMB substrate has an irritant effect on skin and mucosa. In case of possible contact, wash eyes with an abundant volume of water and skin with soap and abundant water. Wash contaminated objects before reusing them. If inhaled, take the person to open air.
20. Chemicals and prepared or used reagents have to be treated as hazardous waste according to the national biohazard safety guideline or regulation.
21. For information on hazardous substances included in the kit please refer to Material Safety Data Sheets. Material Safety Data Sheets for this product are available upon request directly from DRG.

4 REAGENTS

4.1 Reagents provided

1. **Microtiterwells**, 12 x 8 (break apart) strips, 96 wells;
Wells coated with (1-42) β -Amyloid.
2. **Standard (Standard 1-6)**, 6 vials (lyophilized), 0.1 mL;
Concentrations: 1.0 – 5.0 – 15.0 – 30.0 – 60.0 – 120 RU
see „Reagent Preparation“;
contain mercury free preservative.
3. **Control Low & High**, 2 vials (lyophilized), 0.1 mL,
see „Reagent Preparation“
For control values and ranges please refer to vial label or QC-Datasheet.
Contains mercury free preservative.
4. **Sample Diluent**, 1 vial, 25 mL, ready to use,
contains mercury free preservative.
5. **Assay Buffer**, 1 vial, 25 mL, ready to use
contains mercury free preservative.
6. **Enzyme Conjugate**, 1 vial, 20 mL, ready to use,
goat anti human IgG conjugated to horseradish peroxidase;
contains mercury free preservative.
7. **Substrate Solution**, 1 vial, 25 mL, ready to use,
Tetramethylbenzidine (TMB).
8. **Stop Solution**, 1 vial, 14 mL, ready to use,
contains 0.5M H₂SO₄,
Avoid contact with the stop solution. It may cause skin irritations and burns.
9. **Wash Solution**, 1 vial, 30 mL (40X concentrated),
see „Preparation of Reagents“.

Note: Additional *Sample Diluent* for sample dilution is available upon request.

4.2 Materials required but not provided

- A microtiter plate calibrated reader (450 \pm 10 nm) (e.g. the DRG Instruments Microtiter Plate Reader).
- Calibrated variable precision micropipettes.
- Absorbent paper.
- Distilled or deionized water
- Timer
- Semi logarithmic graph paper or software for data reduction

4.3 Storage Conditions

When stored at 2 °C to 8 °C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date.

Opened reagents must be stored at 2 °C to 8 °C. Microtiter wells must be stored at 2 °C to 8 °C. Once the foil bag has been opened, care should be taken to close it tightly again.

Opened kit components (with exception of reconstituted standard) retain activity for two months if stored as described above.

4.4 Reagent Preparation

Bring all reagents and required number of strips to room temperature prior to use.

Standards and Control

Reconstitute the lyophilized contents of each vial with 100 μ L *Sample Diluent*.

Note: *The reconstituted standards and control are stable for one day at 2 °C to 8 °C.*

The reconstituted standards and control have to be diluted 1 + 10 with *Sample Diluent*;
e.g. 10 μ L *Standard or Control* + 100 μ L *Sample Diluent*

This prediluted standards and control have to be used immediately. Remaining volumes have to be discarded.

Wash Solution

Add deionized water to the 40X concentrated Wash Solution.

Dilute 30 mL of concentrated *Wash Solution* with 1170 mL deionized water to a final volume of 1200 mL.

The diluted Wash Solution is stable for 2 weeks at room temperature.

4.5 Disposal of the Kit

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Material Safety Data Sheet.

4.6 Damaged Test Kits

In case of any severe damage to the test kit or components, DRG has to be informed in writing, at the latest, one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

5 SPECIMEN COLLECTION AND PREPARATION

Serum can be used in this assay.

Do not use haemolytic, icteric or lipaemic specimens.

Please note: Samples containing sodium azide should not be used in the assay.

5.1 Specimen Collection

Serum:

Collect blood by venipuncture (e.g. Sarstedt Monovette for serum), allow to clot, and separate serum by centrifugation at room temperature. Do not centrifuge before complete clotting has occurred. Patients receiving anticoagulant therapy may require increased clotting time.

5.2 Specimen Storage and Preparation

Specimens should be capped and may be stored for up to 5 days at 2 °C to 8 °C prior to assaying.

Specimens held for a longer time (up to 12 months) should be frozen only once at -20 °C prior to assay.

Thawed samples should be inverted several times prior to testing.

5.3 Specimen Dilution

Prior to assay the samples have to be diluted 1 + 10 with *Sample Diluent*;

e.g. 10 μ L sample + 100 μ L *Sample Diluent* (mix thoroughly)

Prediluted samples have to be used immediately. Remaining volumes have to be discarded.

If in an initial assay, a specimen is found to contain more than the highest standard, the specimens can be diluted with *Sample Diluent* and reassayed as described in Assay Procedure.

For the calculation of the concentrations this dilution factor has to be taken into account.

Example:

dilution 1:10: 10 μ L diluted sample + 90 μ L *Sample Diluent* (mix thoroughly)

6 ASSAY PROCEDURE

6.1 General Remarks

- All reagents and specimens must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination.
- Absorbance is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.

6.2 Test Procedure

Each run must include a standard curve.

1. Secure the desired number of Microtiter wells in the frame holder.
2. Dispense **20 μ L** of each prediluted **Standard, Control** and prediluted **samples** with new disposable tips into appropriate wells. (See "Reagent Preparation" and "Specimen Dilution")
3. Dispense **200 μ L Assay Buffer** into each well.
4. Incubate for **30 minutes** on a shaker (300-500 rpm) at room temperature.
5. Briskly shake out the contents of the wells.
Rinse the wells **4 times** with diluted **Wash Solution** (400 μ L per well). Strike the wells sharply on absorbent paper to remove residual droplets.
Important note:
The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!
6. Dispense **150 μ L Enzyme Conjugate** into each well.
7. Incubate for **30 minutes** on a shaker (300-500 rpm) at room temperature.
8. Briskly shake out the contents of the wells.
Rinse the wells **4 times** with diluted **Wash Solution** (400 μ L per well). Strike the wells sharply on absorbent paper to remove residual droplets.
9. Add **150 μ L of Substrate Solution** to each well.
10. Incubate for **10 minutes** at room temperature.
11. Stop the enzymatic reaction by adding **100 μ L of Stop Solution** to each well.
12. Determine the absorbance (OD) of each well at **450 \pm 10 nm** with a microtiter plate reader.
It is recommended that the wells be read **within 10 minutes** after adding the *Stop Solution*.

6.3 Calculation of Results

1. Calculate the average absorbance values for each set of standards, controls and patient samples.
2. Using semi-logarithmic graph paper, construct a standard curve by plotting the mean absorbance obtained from each standard against its concentration with absorbance value on the vertical (Y) axis and concentration on the horizontal (X) axis.
3. Using the mean absorbance value for each sample determine the corresponding concentration from the standard curve.
4. Automated method: The results in the IFU have been calculated automatically using a 4 PL (4 Parameter Logistics) curve fit. 4 Parameter Logistics is the preferred method. Other data reduction functions may give slightly different results.
5. The concentration of the samples can be read directly from this standard curve. Samples with concentrations higher than that of the highest standard have to be further diluted or reported as > 120 RU. For the calculation of the concentrations this dilution factor has to be taken into account.

6.3.1 Example of Typical Standard Curve

The following data is for demonstration only and **cannot** be used in place of data generations at the time of assay.

Standard	Optical Units (450 nm)
Standard 1 (1 RU)	0.22
Standard 2 (5 RU)	0.27
Standard 3 (15 RU)	0.41
Standard 4 (30 RU)	0.63
Standard 5 (60 RU)	1.2
Standard 6 (120 RU)	2.5

7 EXPECTED NORMAL VALUES

It is strongly recommended that each laboratory should determine its own normal and abnormal values.

In a study conducted with apparently normal healthy adults, using the DRG Amyloid beta Auto-Ab ELISA the following values are observed:

Population	5 – 95% Percentile	Mean
Men or women, younger or older than 50 years	9.49 - 71.01 RU	21.45 RU

The results alone should not be the only reason for any therapeutic consequences. The results should be correlated to other clinical observations and diagnostic tests.

8 QUALITY CONTROL

Good laboratory practice requires that controls be run with each calibration curve. A statistically significant number of controls should be assayed to establish mean values and acceptable ranges to assure proper performance.

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results. Use controls at both normal and pathological levels.

The controls and the corresponding results of the QC-Laboratory are stated in the QC certificate added to the kit. The values and ranges stated on the QC sheet always refer to the current kit lot and should be used for direct comparison of the results.

It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results.

Employ appropriate statistical methods for analysing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials patient results should be considered invalid. In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.

After checking the above mentioned items without finding any error contact your distributor or DRG directly.

9 PERFORMANCE CHARACTERISTICS

9.1 Assay Dynamic Range

The range of the assay is between 1.7 – 120 RU.

9.2 Specificity of Antibodies (Cross Reactivity)

The autoantibodies recognize the synthetic 1-42 amyloid beta peptide.

9.3 Sensitivity

The analytical sensitivity of the DRG ELISA was calculated by adding 2 standard deviations to the mean of 20 replicate analyses of the Zero Standard (S0) and was found to be 1.7 RU.

9.4 Reproducibility

9.4.1 Intra Assay

The within assay variability is shown below:

Sample	n	Mean (RU)	CV (%)
1	20	32.3	6.2
2	20	50.1	4.7

9.4.2 Inter Assay

The between assay variability is shown below:

Sample	n	Mean (RU)	CV (%)
1	15	22.6	13.4
2	15	35.1	5.88

9.5 Recovery

Diluted Samples have been spiked by adding solutions with known concentrations of autoantibodies against A β in a 1:1 ratio.

The expected values were calculated by addition of half of the values determined for the undiluted samples and half of the values of the known solutions. The % Recovery has been calculated by multiplication of the ratio of the measurements and the expected values with 100.

	Sample 1	Sample 2
Concentration [RU]	37.1	14.5
Average Recovery	91.5	88.6
Range of Recovery [%]	from	87.0
	to	95.7
		85.9
		93.7

9.6 Linearity

	Sample 1	Sample 2
Concentration [RU]	37.1	28.8
Average Recovery	100.8	109.0
Range of Recovery [%]	from	94.3
	to	110.0
		100.7
		113.9

10 LIMITATIONS OF USE

Reliable and reproducible results will be obtained when the assay procedure is performed with a complete understanding of the package insert instruction and with adherence to good laboratory practice. Any improper handling of samples or modification of this test might influence the results.

10.1 Drug Interferences

Until today no substances (drugs) are known to us, which have an influence to the measurement of autoantibodies against A β in a sample.

10.2 High-Dose-Hook Effect

No hook effect was observed in this test up to 480 RU.

11 LEGAL ASPECTS

11.1 Reliability of Results

The test must be performed exactly as per the manufacturer's instructions for use. Moreover the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable national standards and/or laws. This is especially relevant for the use of control reagents. It is important to always include, within the test procedure, a sufficient number of controls for validating the accuracy and precision of the test. The test results are valid only if all controls are within the specified ranges and if all other test parameters are also within the given assay specifications. In case of any doubt or concern please contact DRG.

11.2 Therapeutic Consequences

Therapeutic consequences should never be based on laboratory results alone even if all test results are in agreement with the items as stated under point 11.1. Any laboratory result is only a part of the total clinical picture of a patient.

Only in cases where the laboratory results are in acceptable agreement with the overall clinical picture of the patient should therapeutic consequences be derived.

The test result itself should never be the sole determinant for deriving any therapeutic consequences.

11.3 Liability











Any modification of the test kit and/or exchange or mixture of any components of different lots from one test kit to another could negatively affect the intended results and validity of the overall test. Such modification and/or exchanges invalidate any claim for replacement.

Claims submitted due to customer misinterpretation of laboratory results subject to point 11.2. are also invalid. Regardless, in the event of any claim, the manufacturer's liability is not to exceed the value of the test kit. Any damage caused to the test kit during transportation is not subject to the liability of the manufacturer.

12 REFERENCES / LITERATURE

1. Sambamurti, K., et al., *Neuromolecular Med.*, 1, 1-31 (2002).
2. Lorenzo, A., et al., *Nat. Neurosci.*, 3, 460-464 (2000).

SYMBOLS USED WITH DRG ELISAS

Symbol	English	Deutsch	Français	Español	Italiano
	European Conformity	CE-Konformitätskennzeichnung	Conforme aux normes européennes	Conformidad europea	Conformità europea
	Consult instructions for use	Gebrauchsanweisung beachten	Consulter les instructions d'utilisation	Consulte las Instrucciones	Consultare le istruzioni per l'uso
	In vitro diagnostic device	In-vitro-Diagnostikum	Usage Diagnostic in vitro	Para uso Diagnóstico in vitro	Per uso Diagnostica in vitro
	For research use only	Nur für Forschungszwecke	Seulement dans le cadre de recherches	Sólo para uso en investigación	Solo a scopo di ricerca
	Catalogue number	Katalog-Nr.	Référence	Número de catálogo	No. di Cat.
	Lot. No. / Batch code	Chargen-Nr.	No. de lot	Número de lote	Lotto no
	Contains sufficient for <n> tests/	Ausreichend für "n" Ansätze	Contenu suffisant pour "n" tests	Contenido suficiente para <n> ensayos	Contenuto sufficiente per "n" saggi
	Storage Temperature	Lagerungstemperatur	Température de conservation	Temperatura de conservación	Temperatura di conservazione
	Expiration Date	Mindesthaltbarkeitsdatum	Date limite d'utilisation	Fecha de caducidad	Data di scadenza
	Legal Manufacturer	Hersteller	Fabricant	Fabricante	Fabbricante
<i>Distributed by</i>	Distributor	Vertreiber	Distributeur	Distribuidor	Distributore
<i>Content</i>	Content	Inhalt	Contenu	Contenido	Contenuto
<i>Volume/No.</i>	Volume / No.	Volumen/Anzahl	Volume/Numéro	Volumen/Número	Volume/Quantità
<i>Microtiterwells</i>	Microtiterwells	Mikrotiterwells	Plaques de micro-titration	Placas multipocillo	Micropozzetti
<i>Antiserum</i>	Antiserum	Antiserum	Antisérum	Antisero	Antisiero
<i>Enzyme Conjugate</i>	Enzyme Conjugate	Enzymkonjugat	Conjugué enzymatique	Conjugado enzimático	Tracciante enzimatico
<i>Enzyme Complex</i>	Enzyme Complex	Enzymkomplex	Complexe enzymatique	Complex enzimático	Complesso enzimatico
<i>Substrate Solution</i>	Substrate Solution	Substratlösung	Solution substrat	Solución de sustrato	Soluzione di substrato
<i>Stop Solution</i>	Stop Solution	Stopplösung	Solution d'arrêt	Solución de parada	Soluzione d'arresto
<i>Zero Standard</i>	Zero Standard	Nullstandard	Zero Standard	Estándar cero	Standard zero
<i>Standard</i>	Standard	Standard	Standard	Estándar	Standard
<i>Control</i>	Control	Kontrolle	Contrôle	Control	Controllo
<i>Assay Buffer</i>	Assay Buffer	Assaypuffer	Tampon d'essai	Tampón de ensayo	Tampone del test
<i>Wash Solution</i>	Wash Solution	Waschlösung	Solution de lavage	Solución de lavado	Soluzione di lavaggio
<i>1N NaOH</i>	1N NaOH	1N NaOH	1N NaOH	1N NaOH	1N NaOH (idrossido di sodio 1N)
<i>1 N HCl</i>	1 N HCl	1 N HCl	1N HCl	1 N HCl	
<i>Sample Diluent</i>	Sample Diluent	Probenverdünnungsmedium	Solution pour dilution de l'échantillon	Solución para dilución de la muestra	Diluyente dei campioni
<i>Conjugate Diluent</i>	Conjugate Diluent	Konjugatverdünnungsmedium	Solution pour dilution du conjugué	Solución para dilución del conjugado	Diluyente del tracciante